

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 06/14/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:11
INTERNAL STANDARD	VERIFIED	16:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:12
BLANK TEST	0.00	16:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:13
BLANK TEST	0.00	16:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:14
BLANK TEST	0.00	16:15

