

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 06/23/2020

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:13
INTERNAL STANDARD	VERIFIED	12:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:14
BLANK TEST	0.00	12:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:15
BLANK TEST	0.00	12:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:16
BLANK TEST	0.00	12:17

