

SOLUTION CHANGE

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 06/03/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:58
INTERNAL STANDARD	VERIFIED	15:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:59
BLANK TEST	0.00	15:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:01
BLANK TEST	0.00	16:02

