

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 05/14/2020

OPERATOR NAME:
KURT M MULLIGAN
Permit #: 44223
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:15
INTERNAL STANDARD	VERIFIED	09:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:17
BLANK TEST	0.00	09:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19

