

SUPERVISOR TEST

-----  
STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108  
-----

Date: 05/23/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:38
INTERNAL STANDARD	VERIFIED	23:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:39
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42

