

SOLUTION CHANGE

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 05/04/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:02
INTERNAL STANDARD	VERIFIED	12:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:02
BLANK TEST	0.00	12:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:03
BLANK TEST	0.00	12:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:05
BLANK TEST	0.00	12:05

