

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 07/22/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:11
INTERNAL STANDARD	VERIFIED	21:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:12
BLANK TEST	0.00	21:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:13
BLANK TEST	0.00	21:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:14
BLANK TEST	0.00	21:15

