

SOLUTION CHANGE

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 07/03/2020

OPERATOR NAME:  
CHRISTOPHER R JORDAN  
Permit #: 46293  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:17
INTERNAL STANDARD	VERIFIED	00:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:21

