

SOLUTION CHANGE

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 08/02/2020

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:36
INTERNAL STANDARD	VERIFIED	15:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:37
BLANK TEST	0.00	15:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:38
BLANK TEST	0.00	15:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:39
BLANK TEST	0.00	15:40

