

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 08/26/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:06
INTERNAL STANDARD	VERIFIED	16:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:06
BLANK TEST	0.00	16:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:07
BLANK TEST	0.00	16:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:09
BLANK TEST	0.00	16:10

