

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 08/08/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:19
INTERNAL STANDARD	VERIFIED	10:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:20
BLANK TEST	0.00	10:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:21
BLANK TEST	0.00	10:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:22
BLANK TEST	0.00	10:23

