

SUPERVISOR TEST

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 09/13/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:20
INTERNAL STANDARD	VERIFIED	11:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:20
BLANK TEST	0.00	11:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:22
BLANK TEST	0.00	11:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:23
BLANK TEST	0.00	11:24

