

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 09/20/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0:00	16:21
INTERNAL STANDARD	VERIFIED	16:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:21
BLANK TEST	0.00	16:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:22
BLANK TEST	0.00	16:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:24
BLANK TEST	0.00	16:24

