

SUPERVISOR TEST

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 09/08/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:00
INTERNAL STANDARD	VERIFIED	16:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:02
BLANK TEST	0.00	16:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:03
BLANK TEST	0.00	16:04

