

SOLUTION CHANGE

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 10/01/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:54
INTERNAL STANDARD	VERIFIED	22:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:55
BLANK TEST	0.00	22:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:56
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:57
BLANK TEST	0.00	22:58

