

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 10/30/2020

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:06
INTERNAL STANDARD	VERIFIED	16:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:06
BLANK TEST	0.00	16:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:07
BLANK TEST	0.00	16:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:09
BLANK TEST	0.00	16:09

