



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of APPLIED CONCEPTS INC, STALKER II MDR RADAR, serial number AS012743, performed on October 22, 2020 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection /maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under
ESRA by Joshua Vinehout
on 10/28/2020 9:52 AM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**RADAR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Stony Point Town Police Department
Manufacturer: APPLIED CONCEPTS INC Model: STALKER II MDR
Date: October 22, 2020 Serial Number: AS012743

I hereby certify that APPLIED CONCEPTS INC RADAR, model STALKER II MDR, serial number AS012743// N/A // N/A, has been calibrated using standards whose accuracies are established by the National Bureau of Standards, or have been derived by the ratio type of self calibration techniques. Calibration has been effected by controlled tests performed on the date indicated above.

Test No	Internal Standard	Test Result
1	PAS MPH	PAS MPH
Test No	Certification Standard (Stationary)	Test Result
1	40 MPH	40 MPH
2	25 MPH	25 MPH
Test No	Certification Standard (Moving)	Test Result
1	40 MPH	15 MPH
	25 MPH	

Turning Fork Certification / Date of Certification: October 22, 2020


Serial Number	Frequency	Test Result
FB341617	4163 HZ	40 MPH

The above stated tuning fork has been tested and found to oscillate at 4163 Hertz. It will cause a calibration signal of 40 MPH when used with a Doppler traffic radar operating at 34,700 Mhz.

Serial Number	Frequency	Test Result
FA234347	2615 HZ	25 MPH

The above stated tuning fork has been tested and found to oscillate at 2615 Hertz. It will cause a calibration signal of 25 MPH when used with a Doppler traffic radar operating at 34,700 Mhz.

I further certify that the entries made in these records were made at the time that the inspection /maintenance/calibration of the above identified RADAR was performed, or within a reasonable time thereafter.



Digitally signed under ESRA
by Joseph Iavarone
on 10/22/2020 12:32 PM

Joseph Iavarone
Highway Safety Equipment Technician
Office of Public Safety

**RADAR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Stony Point Town Police Department
Manufacturer: APPLIED CONCEPTS INC
Model: STALKER II MDR
Serial Number: AS012743
Date: October 22, 2020

- 1. General appearance and condition of this device -----: OK**

- 2. Turn system ON -----: OK**

- 3. Check visual displays/indicators -----: OK**

- 4. System cycles through all modes -----: OK**

- 5. Record any repairs made to this instrument and list services provided:
VERIFIED CALIBRATION**



Joseph Iavarone

Digitally signed under ESRA
by Joseph Iavarone
on 10/22/2020 12:32 PM

Joseph Iavarone
Highway Safety Equipment Technician
Office of Public Safety