

SOLUTION CHANGE

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 12/01/2020

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:22
INTERNAL STANDARD	VERIFIED	12:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:22
BLANK TEST	0.00	12:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:24
BLANK TEST	0.00	12:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:25
BLANK TEST	0.00	12:26

