

SUPERVISOR TEST

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 12/23/2020

OPERATOR NAME:  
CHRISTOPHER R JORDAN  
Permit #: 46293  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:28
INTERNAL STANDARD	VERIFIED	09:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32

