

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 12/06/2020

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 16:27 |
| INTERNAL STANDARD | VERIFIED | 16:27 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 16:27 |
| BLANK TEST | 0.00 | 16:28 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 16:28 |
| BLANK TEST | 0.00 | 16:29 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 16:30 |
| BLANK TEST | 0.00 | 16:30 |

