

SUPERVISOR TEST

 STONY POINT
 TOWN POLICE DEPT.
 DATAMASTER dmt: 112108

Date: 01/08/2022

OPERATOR NAME:
 CHRISTOPHER R JORDAN
 Permit #: 46293
 Operator Agency/Dept: STONY POINT TOWN PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:31
INTERNAL STANDARD	VERIFIED	16:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:32
BLANK TEST	0.00	16:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:33
BLANK TEST	0.00	16:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:34
BLANK TEST	0.00	16:35

