

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 02/13/2022

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:22
INTERNAL STANDARD	VERIFIED	16:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:23
BLANK TEST	0.00	16:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:24
BLANK TEST	0.00	16:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:25
BLANK TEST	0.00	16:26

