

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 02/23/2022

OPERATOR NAME:
MATHEW J GAYNOR
Permit #: 46292
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10

