

SUPERVISOR TEST

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 01/26/2022

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:45
INTERNAL STANDARD	VERIFIED	20:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:45
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:46
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:47
BLANK TEST	0.00	20:48

