

SUPERVISOR TEST

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 01/30/2022

OPERATOR NAME:  
CHRISTOPHER R JORDAN  
Permit #: 46293  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:52
INTERNAL STANDARD	VERIFIED	15:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:53
BLANK TEST	0.00	15:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:54
BLANK TEST	0.00	15:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:55
BLANK TEST	0.00	15:56

