

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 03/14/2022

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:11
INTERNAL STANDARD	VERIFIED	10:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:12
BLANK TEST	0.00	10:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:14
BLANK TEST	0.00	10:15

