

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 03/20/2022

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:40
INTERNAL STANDARD	VERIFIED	17:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:40
BLANK TEST	0.00	17:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:43
BLANK TEST	0.00	17:44

