



10B Airline Drive, Albany, New York 12235
518-457-3146 (P)
518-457-2552 (F)
www.weighsyr.ny.gov

Bureau of Weights & Measures

Wheel-Load Weigher Test Report

Test No: W23-015

For: Stony Point Police Department
79 Route 210
Stony Poin NY 10980

The scales listed below have been compared to Standards of New York State which are traceable to the National Institute of Standards and Technology (NIST). Tolerances are 2% of the applied test load. Scales sets are tested as a pair and must be used as such.

Set	S1	Status: PASS	Tested:	3/8/2023		
	Scale MFR/Model: Haenni		WL101	S/Ns:	24522	24523
Load/Reading	9,270 9,350	19,160 19,275		28,830 29,050	38,940 39,175	
Set	S2	Status: PASS	Tested:	3/8/2023		
	Scale MFR/Model: Haenni		wl101	S/Ns:	24524	24525
Load/Reading	9,130 9,150	19,210 19,250		30,040 30,150	38,920 39,025	
Set	S3	Status: PASS	Tested:	3/8/2023		
	Scale MFR/Model: Haenni		WL101	S/Ns:	24526	24527
Load/Reading	9,320 9,325	19,330 19,250		29,100 29,000	39,600 39,575	

Weights & Measures Officials

Enforcement Officials

David J. Singh

SGT J. Wood #657

[Signature]
